

SUPERVISOR TEST

WARREN COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 144506

Date: 06/16/2025

OPERATOR NAME:  
 MICHAEL A CAMPBELL  
 Permit #: 42422  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:20
INTERNAL STANDARD	VERIFIED	23:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:21
BLANK TEST	0.00	23:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:22
BLANK TEST	0.00	23:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:23
BLANK TEST	0.00	23:24

