

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 06/20/2025

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:21
INTERNAL STANDARD	VERIFIED	20:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:21
BLANK TEST	0.00	20:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:22
BLANK TEST	0.00	20:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:24
BLANK TEST	0.00	20:24

[Handwritten signature]
#870

