

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 06/28/2025

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:51
INTERNAL STANDARD	VERIFIED	15:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:52
BLANK TEST	0.00	15:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:53
BLANK TEST	0.00	15:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:54
BLANK TEST	0.00	15:55

