

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 07/05/2025

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:40
INTERNAL STANDARD	VERIFIED	02:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:41
BLANK TEST	0.00	02:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:42
BLANK TEST	0.00	02:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:43
BLANK TEST	0.00	02:44

[Handwritten signature]
#870

