

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 07/13/2025

OPERATOR NAME:  
STEVEN T BUNIO  
Permit #: 44983  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:55
INTERNAL STANDARD	VERIFIED	12:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:55
BLANK TEST	0.00	12:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:57
BLANK TEST	0.00	12:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:58
BLANK TEST	0.00	12:59

