

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 07/26/2025

OPERATOR NAME:
JEREMY M CRISPELL
Permit #: 51155
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:26
INTERNAL STANDARD	VERIFIED	09:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:27
BLANK TEST	0.00	09:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:28
BLANK TEST	0.00	09:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:29
BLANK TEST	0.00	09:30

Handwritten signature: JMW

