

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 08/15/2025

OPERATOR NAME:
JESSE P SUMMO
Permit #: 47506
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:44
INTERNAL STANDARD	VERIFIED	19:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	19:45
BLANK TEST	0.00	19:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	19:46
BLANK TEST	0.00	19:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:47
BLANK TEST	0.00	19:48

JS 174

