

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 08/22/2025

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:30
INTERNAL STANDARD	VERIFIED	13:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:31
BLANK TEST	0.00	13:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:32
BLANK TEST	0.00	13:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:33
BLANK TEST	0.00	13:34

