

SUPERVISOR TEST




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 08/29/2025

OPERATOR NAME:
JESSE P SUMMO
Permit #: 47506
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:29
INTERNAL STANDARD	VERIFIED	22:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:30
BLANK TEST	0.00	22:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:31
BLANK TEST	0.00	22:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:32
BLANK TEST	0.00	22:33



Jesse P Summo 174