

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 09/19/2025

OPERATOR NAME:  
J RIEHL  
Permit #: 47503  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:07
INTERNAL STANDARD	VERIFIED	03:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:07
BLANK TEST	0.00	03:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:08
BLANK TEST	0.00	03:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:10
BLANK TEST	0.00	03:11

*#820*

