

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 09/28/2025

OPERATOR NAME:
MICHAEL A CAMPBELL
Permit #: 42422
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:01
INTERNAL STANDARD	VERIFIED	23:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:01
BLANK TEST	0.00	23:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:02
BLANK TEST	0.00	23:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:03
BLANK TEST	0.00	23:04

