

SOLUTION CHANGE

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 10/01/2025

OPERATOR NAME:
MICHAEL A CAMPBELL
Permit #: 42422
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:54
INTERNAL STANDARD	VERIFIED	16:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:55
BLANK TEST	0.00	16:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:56
BLANK TEST	0.00	16:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:57
BLANK TEST	0.00	16:58

