

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 10/03/2025

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:57
INTERNAL STANDARD	VERIFIED	09:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:58
BLANK TEST	0.00	09:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:59
BLANK TEST	0.00	10:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:00
BLANK TEST	0.00	10:01

PT Bunio 10-3-25

