

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 10/11/2025

OPERATOR NAME:  
J RIEHL  
Permit #: 47503  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:41
INTERNAL STANDARD	VERIFIED	03:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:42
BLANK TEST	0.00	03:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:43
BLANK TEST	0.00	03:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:44
BLANK TEST	0.00	03:45



#822

