

SUPERVISOR TEST




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 10/17/2025

OPERATOR NAME:
ETHAN P OSTAPCZUK
Permit #: 50075
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:42
INTERNAL STANDARD	VERIFIED	06:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:43
BLANK TEST	0.00	06:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:44
BLANK TEST	0.00	06:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:45
BLANK TEST	0.00	06:46



UPB #152