

SOLUTION CHANGE

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 11/01/2025

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:14
INTERNAL STANDARD	VERIFIED	11:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:15
BLANK TEST	0.00	11:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:16
BLANK TEST	0.00	11:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:17
BLANK TEST	0.00	11:18

