

SUPERVISOR TEST

WARREN COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 11/08/2025

OPERATOR NAME:
 J RIEHL
 Permit #: 47503
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:16
INTERNAL STANDARD	VERIFIED	01:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:17
BLANK TEST	0.00	01:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:18
BLANK TEST	0.00	01:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:19
BLANK TEST	0.00	01:20

[Handwritten signature]
 #820

