

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506 - Warrensburg Station

Date: 11/21/2025

OPERATOR NAME:
ANTHONY E CARDENAS
Permit #: 50642
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:36
INTERNAL STANDARD	VERIFIED	21:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	21:36
BLANK TEST	0.00	21:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	21:37
BLANK TEST	0.00	21:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	21:39
BLANK TEST	0.00	21:39

[Signature] 184
3

