

SOLUTION CHANGE




WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 12/01/2025

OPERATOR NAME:  
ETHAN P OSTAPCZUK  
Permit #: 50075  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:32
INTERNAL STANDARD	VERIFIED	22:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:33
BLANK TEST	0.00	22:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:34
BLANK TEST	0.00	22:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:35
BLANK TEST	0.00	22:36



*[Handwritten signature]* 152