

SUPERVISOR TEST




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 12/07/2025

OPERATOR NAME:
JESSE P SUMMO
Permit #: 47506
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:29
INTERNAL STANDARD	VERIFIED	01:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:30
BLANK TEST	0.00	01:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:31
BLANK TEST	0.00	01:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:32
BLANK TEST	0.00	01:33



Jesse P Summo 174