

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 12/20/2025

OPERATOR NAME:
MICHAEL A CAMPBELL
Permit #: 42422
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:41
INTERNAL STANDARD	VERIFIED	00:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:41
BLANK TEST	0.00	00:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:42
BLANK TEST	0.00	00:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:44
BLANK TEST	0.00	00:44

