

SUPERVISOR TEST




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 12/26/2025

OPERATOR NAME:
ETHAN P OSTAPCZUK
Permit #: 50075
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:08
INTERNAL STANDARD	VERIFIED	00:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:08
BLANK TEST	0.00	00:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:09
BLANK TEST	0.00	00:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:11
BLANK TEST	0.00	00:11



Ethan P. Ostapczuk 152