

SOLUTION CHANGE

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 01/02/2026

OPERATOR NAME:  
J RIEHL  
Permit #: 47503  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:11
INTERNAL STANDARD	VERIFIED	23:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:11
BLANK TEST	0.00	23:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:12
BLANK TEST	0.00	23:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:14
BLANK TEST	0.00	23:15

*J Riehl*  
#820

