

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 01/09/2026

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:34
INTERNAL STANDARD	VERIFIED	17:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:35
BLANK TEST	0.00	17:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:36
BLANK TEST	0.00	17:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:37
BLANK TEST	0.00	17:38

PT Steve Bunio #789

