

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 01/23/2026

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:26
INTERNAL STANDARD	VERIFIED	00:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:26
BLANK TEST	0.00	00:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:27
BLANK TEST	0.00	00:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:29
BLANK TEST	0.00	00:29

[Handwritten signature]
#870

