

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 02/03/2026

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:17
INTERNAL STANDARD	VERIFIED	17:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:18
BLANK TEST	0.00	17:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:19
BLANK TEST	0.00	17:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:20
BLANK TEST	0.00	17:21

Steven Bunio #789

