

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 02/06/2026

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:34
INTERNAL STANDARD	VERIFIED	01:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.08	01:35
BLANK TEST	0.00	01:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.08	01:36
BLANK TEST	0.00	01:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.08	01:37
BLANK TEST	0.00	01:38

[Handwritten signature]
#820

