

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 02/10/2026

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:22
INTERNAL STANDARD	VERIFIED	14:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:23

Blank Error

[Handwritten signature]
#820

