

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 02/13/2026

OPERATOR NAME:
J T COON
Permit #: 36043
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:38
INTERNAL STANDARD	VERIFIED	02:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:39
BLANK TEST	0.00	02:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:40

Blank Error

