

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 02/13/2026

OPERATOR NAME:  
ROBERT A ROSE  
Permit #: 51175  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:44
INTERNAL STANDARD	VERIFIED	06:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:44
BLANK TEST	0.00	06:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:45

Blank Error

*RRR #897*

