

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 02/18/2026

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:45
INTERNAL STANDARD	VERIFIED	20:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:45
BLANK TEST	0.00	20:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:47
BLANK TEST	0.00	20:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:48
BLANK TEST	0.00	20:49

[Handwritten signature]
#820

