

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 02/20/2026

OPERATOR NAME:
JESSE P SUMMO
Permit #: 47506
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:53
INTERNAL STANDARD	VERIFIED	06:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:54
BLANK TEST	0.00	06:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:55
BLANK TEST	0.00	06:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:56
BLANK TEST	0.00	06:57

JS 174

