

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 02/27/2026

OPERATOR NAME:
RYAN C SCHROECK
Permit #: 41316
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:14
INTERNAL STANDARD	VERIFIED	12:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	12:14
BLANK TEST	0.00	12:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	12:16
BLANK TEST	0.00	12:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:17
BLANK TEST	0.00	12:18

pu/1709

