

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 01/09/2026

OPERATOR NAME:  
STEVEN T BUNIO  
Permit #: 44983  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 25330\

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:10
INTERNAL STANDARD	VERIFIED	07:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:11
BLANK TEST	0.00	07:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:12
BLANK TEST	0.00	07:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:13
BLANK TEST	0.00	07:14

*Handwritten signature and permit number: #44983*

