

SOLUTION CHANGE

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 02/03/2026

OPERATOR NAME:  
STEVEN T BUNIO  
Permit #: 44983  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:53
INTERNAL STANDARD	VERIFIED	11:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:53
BLANK TEST	0.00	11:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:54
BLANK TEST	0.00	11:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:55
BLANK TEST	0.00	11:56

*PO Steven T Bunio 2/3/26*

