

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 02/27/2026

OPERATOR NAME:  
RYAN C SCHROECK  
Permit #: 41316  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:32
INTERNAL STANDARD	VERIFIED	11:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:32
BLANK TEST	0.00	11:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	11:33
BLANK TEST	0.00	11:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:35
BLANK TEST	0.00	11:36

*Handwritten signature*

