

SOLUTION CHANGE

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 01/02/2026

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:44
INTERNAL STANDARD	VERIFIED	23:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:44
BLANK TEST	0.00	23:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:45
BLANK TEST	0.00	23:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:46
BLANK TEST	0.00	23:47

#800

