

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 01/09/2026

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:56
INTERNAL STANDARD	VERIFIED	08:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:56
BLANK TEST	0.00	08:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:58
BLANK TEST	0.00	08:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:59
BLANK TEST	0.00	09:00

PO Steven Bunio #739

