

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 01/23/2026

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:01
INTERNAL STANDARD	VERIFIED	01:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:02
BLANK TEST	0.00	01:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:03
BLANK TEST	0.00	01:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:04
BLANK TEST	0.00	01:05

#870

