

SOLUTION CHANGE

WARREN COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 02/03/2026

OPERATOR NAME:
 STEVEN T BUNIO
 Permit #: 44983
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:18
INTERNAL STANDARD	VERIFIED	15:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:19
BLANK TEST	0.00	15:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:20
BLANK TEST	0.00	15:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:21
BLANK TEST	0.00	15:22

Handwritten signatures and initials, including a large signature and the number 789.

