

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 02/20/2026

OPERATOR NAME:
JESSE P SUMMO
Permit #: 47506
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:58
INTERNAL STANDARD	VERIFIED	05:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	05:59
BLANK TEST	0.00	06:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:00
BLANK TEST	0.00	06:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:01
BLANK TEST	0.00	06:02

JS 174

