

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 02/27/2026

OPERATOR NAME:
RYAN C SCHROECK
Permit #: 41316
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:34
INTERNAL STANDARD	VERIFIED	12:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:34
BLANK TEST	0.00	12:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:35
BLANK TEST	0.00	12:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:37
BLANK TEST	0.00	12:37

RS/S

