

SOLUTION CHANGE

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 03/01/2026

OPERATOR NAME:
RYAN C SCHROECK
Permit #: 41316
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:40
INTERNAL STANDARD	VERIFIED	12:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:41
BLANK TEST	0.00	12:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:42
BLANK TEST	0.00	12:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:43
BLANK TEST	0.00	12:44

PG 724

