


**BREATH SCREENING DEVICE
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Warren County Sheriff's Office
Manufacturer: INTOXIMETERS INC Model: ALCO-SENSOR FST
Serial Number: 216084
Date: May 07, 2020

Reference Solution Lot Number: 19281
Reference Solution Standard: 0.100 % at 34.00 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.097 %	34.02 degrees centigrade
2	0.097 %	34.02 degrees centigrade
3	0.098 %	34.02 degrees centigrade
4	0.097 %	34.01 degrees centigrade



Michael J. Canary
Highway Safety Equipment Technician
Office of Public Safety

**BREATH SCREENING DEVICE
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Warren County Sheriff's Office
Manufacturer: INTOXIMETERS INC
Model: ALCO-SENSOR FST
Serial Number: 216084
Date: May 07, 2020

1. General Appearance and condition of this device -----: OK

2. Check battery indicators -----: OK

3. Check zero flashing -----: OK

**4. Record any repairs made to this instrument and list services provided: VERIFIED
CALIBRATION**



Michael J. Canary
Highway Safety Equipment Technician
Office of Public Safety

