



**CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS
MAINTAINED IN THE REGULAR COURSE OF BUSINESS**

I, Frank Sgarlata, Highway Safety Equipment Technician 2, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of GUTH LABORATORIES, 12V500 SIMULATOR, serial number MP4757, performed on July 31, 2025 by Office of Public Safety employee Martin Schilling, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/ maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Frank Sgarlata".

Digitally signed under ESRA
by Frank Sgarlata
on 08/01/2025 02:49 PM

**Frank Sgarlata
Highway Safety Equipment Technician 2
Office of Public Safety**

**SIMULATOR
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Warren County Sheriff's Office

Manufacturer: GUTH LABORATORIES

Model: 12V500

Serial Number: MP4757

Date: July 31, 2025

Breath Simulator Temperature
34.01 degrees centigrade



**Division of Criminal
Justice Services**

I hereby certify that I have performed any necessary maintenance procedures and verified the calibration of GUTH LABORATORIES simulator model 12V500, serial number MP4757, and have determined that it is accurate.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified simulator was performed, or within a reasonable time thereafter.

A handwritten signature in black ink, appearing to read "Martin Schilling".

Digitally signed under ESRA
by Martin Schilling
on 07/31/2025 11:58 AM

**Martin Schilling
Highway Safety Equipment Technician
Office of Public Safety**

**SIMULATOR
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Warren County Sheriff's Office

Manufacturer: GUTH LABORATORIES

Model: 12V500

Serial Number: MP4757

Date: July 31, 2025

1. General appearance and condition of this device -----: OK

2. Check propeller rotation -----: OK

3. Check for air leaks -----: OK

4. Check for proper temperature -----: OK

5. Record any repairs made to this instrument and list services provided : VERIFIED

CALIBRATION



**Division of Criminal
Justice Services**

A handwritten signature in black ink that reads "Martin Schilling".

Digitally signed under ESRA
by Martin Schilling
on 07/31/2025 11:58 AM

**Martin Schilling
Highway Safety Equipment Technician
Office of Public Safety**

NIST Traceable
Calibration Report

REPORT NUMBER

1777015

Reference Number: 180845700

PO Number: MSCHILLING013125

NY State Div Of Criminal Justice Svcs
80 South Swan Street
Albany, NY 12210 United States

Manufacturer: Alpha Technics
Model Number: 91-5443
Description: Temperature, Thermometer
Asset Number: 20BMM01712
Serial Number: 20BMM01712
Procedure: DS Alpha Technics 5000 Series f/NY State DCJS

Calibration Date: 02/24/2025
Calibration Due Date: 02/24/2026
Condition As Found: In Tolerance
Condition As Left: In Tolerance, No adjustment

Remarks:

NIST-traceable calibration performed on the unit referenced above in accordance with customer requirements, published specifications and the lab's standard operating procedures. No adjustments were made to the unit.

Standards Used

Standard ID	Manufacturer	Model Number	Description	Cal Date	Due Date
CP32033	Hart Scientific	5628	Temperature, Probe	7/25/2024	1/31/2026
CP32035	Hart Scientific	5628	Temperature, Probe	12/11/2024	6/30/2026
CP32036	Hart Scientific	5628	Temperature, Probe	4/09/2024	10/31/2025
CP32042	Hart Scientific	5680	Temperature, Probe, SPRT	3/25/2024	12/31/2025
CP371282	Fluke Corporation	1594A	Precision Readout	5/27/2022	3/31/2025

Calibration Data

Function Tested	Nominal / Reference Value	Measured Value	OOT	Calibration Tolerance <i>g: = Guard Banding Applied</i>	TUR	EMU
Temperature	-15.000 °C					
As Found & As Left	-15.000	-14.99		-15.050 to -14.950 °C	4.3:1	± 0.012 °C
	0.000 °C					
As Found & As Left	0.000	0.01		-0.030 to 0.030 °C	4.0:1	± 0.0075 °C
	75.000 °C					
As Found & As Left	75.000	74.99		74.970 to 75.030 °C	3.3:1	± 0.0091 °C
	125.000 °C					
As Found & As Left	125.000	124.98		124.950 to 125.050 °C	3.8:1	± 0.013 °C

Temperature: 20 °C
Humidity: 27 %RH
Rpt. No.: 1777015

Calibration Performed By:		Quality Reviewer:	
Arnel Bautista	Technician, Metrology	James Alexander	2/24/2025
Name	Title	Name	Date

This report may not be reproduced, except in full, without written permission of Innocal. The results stated in this report relate only to the items tested or calibrated. Measurements reported herein are traceable to SI units via national standards maintained by NIST and were performed in compliance with MIL-STD-45662A, ANSI/NCSL Z540-1-1994, 10CFR50, Appendix B, ISO 9002-94, and ISO 17025:2017. Conformance based on Simple Acceptance as defined in ILAC G8 with a < 50% PFA. The estimated measurement uncertainty (EMU), if reported on this certificate, is being reported at a confidence level of 95% or K=2 unless otherwise noted in the remarks section.

SERVICE AUTHORIZATION FORM
Highway Safety Technology Unit
80 South Swan Street
Albany, NY 12210
Ph: 518-485-7636 / 518-402-0689
Fax: 518-457-6869

WJ 31-21
7

THIS FORM IS USED TO AUTHORIZE THE DIVISION OF CRIMINAL JUSTICE SERVICES TO PERFORM SERVICE ON THE INSTRUMENT IDENTIFIED. THIS AUTHORIZATION MUST ACCOMPANY EACH PIECE OF EQUIPMENT WHEN IT IS PRESENTED TO THE EQUIPMENT REPAIR CENTER FOR SERVICE. **PLEASE SIGN AND DATE FORM.**

SECTION I - AGENCY INFORMATION

Name of Agency Warren County Sheriff's Office		Today's Date 07/31/2025
Address 1400 State Route 9	City, State, ZIP Lake George, NY 12845	
Contact Person Patrol Officer James Riehl	Contact Telephone (802) 287-0833	
Contact Email JAMES.RIEHL@WARRENCOUNTYSHERIFFNY.US	Best Way to Reach You During Appointment Period Phone	

SECTION II - INSTRUMENT INFORMATION

Instrument Manufacturer (Please Circle) NPAS DMT <u>GUTH</u> Applied Concepts (Stalker) Decatur (Genesis) Kustom (Eagle/Falcon) MPH Alco-Sensor Other:		
Model Number 12V500	Instrument Serial Number MP4757	
Antenna Serial Number (if applicable) Antenna 1 -- / Antenna 2 --	Type of maintenance (check appropriate) <input checked="" type="checkbox"/> Calibration Check <input type="checkbox"/> Repair	
Brief Description of Malfunction		
List all parts, cables and/or accessories being submitted. Tuning forks must accompany all radar units (if applicable). (Please Circle)		
DMT/SIM: DMT <u>SIM</u> <u>Tubes</u> Keyboard <u>Other: Power Cord</u>	Radar: Control Unit Antenna(s) Antenna Cable(s) Power Cord Forks Other:	
<i>I acknowledge I am signing this Service Authorization Form as an act and deed of said agency, organization, or municipality, and that I am duly authorized to sign same for the uses and purposes mentioned herein. PLEASE SIGN AND DATE.</i>		
Signature 	Date 07/31/2025	

DCJS USE ONLY

Export °C <u>34.00</u> <u>34.00</u> <u>34.00</u>	Digital °C <u>34.01</u> <u>34.01</u> <u>34.01</u>	Thermometer: <u>20 RHM 1212</u> Calibration Due: <u>9-24-06</u> Air Leak Test: <u>OK</u> Propeller Rotation: <u>OK</u>
---	--	--