



**CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS
MAINTAINED IN THE REGULAR COURSE OF BUSINESS**

I, Frank Sgarlata, Highway Safety Equipment Technician 2, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of GUTH LABORATORIES, 12V500 SIMULATOR, serial number MP4824, performed on July 22, 2024 by Office of Public Safety employee Jonathan Mascolo, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/ maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Frank Sgarlata".

Digitally signed under ESRA
by Frank Sgarlata
on 08/16/2024 11:00 AM

**Frank Sgarlata
Highway Safety Equipment Technician 2
Office of Public Safety**

**SIMULATOR
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Warren County Sheriff's Office
Manufacturer: GUTH LABORATORIES
Model: 12V500
Serial Number: MP4824
Date: July 22, 2024

Breath Simulator Temperature
34.02 degrees centigrade



**Division of Criminal
Justice Services**

I hereby certify that I have performed any necessary maintenance procedures and verified the calibration of GUTH LABORATORIES simulator model 12V500, serial number MP4824, and have determined that it is accurate.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified simulator was performed, or within a reasonable time thereafter.

A handwritten signature in black ink that reads "Jonathan Mascolo".

Digitally signed under ESRA
by Jonathan Mascolo
on 08/16/2024 10:54 AM

**Jonathan Mascolo
Highway Safety Equipment Technician
Office of Public Safety**

**SIMULATOR
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Warren County Sheriff's Office

Manufacturer: GUTH LABORATORIES

Model: 12V500

Serial Number: MP4824

Date: July 22, 2024

1. General appearance and condition of this device -----: OK

2. Check propeller rotation -----: OK

3. Check for air leaks -----: OK

4. Check for proper temperature -----: OK

**5. Record any repairs made to this instrument and list services provided : VERIFIED
CALIBRATION**



**Division of Criminal
Justice Services**

Jonathan Mascolo

Digitally signed under ESRA
by Jonathan Mascolo
on 08/16/2024 10:54 AM

**Jonathan Mascolo
Highway Safety Equipment Technician
Office of Public Safety**

625 East Bunker Court
Vernon Hills, Illinois 60061
PH: 866-466-6225
Fax: 847-327-2993
www.innocalsolutions.com

NIST Traceable
Calibration Report

NY State Div Of Criminal Justice Svcs
80 South Swan Street
Albany, NY 12210 United States

REPORT NUMBER
1756976

Reference Number: 164871800
PO Number: MSCHILLING50824

Manufacturer: Ertco
Model Number: 4400
Description: Temperature, Dig thermometer w/probe
Asset Number: 39808
Serial Number: 307443
Procedure: DS Eutechnics 4400-EP642

Calibration Date: 05/31/2024
Calibration Due Date: 05/31/2025
Condition As Found: In Tolerance
Condition As Left: In Tolerance, No adjustment

Remarks:

NIST-traceable calibration performed on the unit referenced above in accordance with customer requirements, published specifications and the lab's standard operating procedures. No adjustments were made to the unit.

Standards Used

Standard ID	Manufacturer	Model Number	Description	Cal Date	Due Date
CP106957	Hart Scientific	5699	Temperature, Probe, SPRT	6/05/2023	12/31/2024
CP32034	Hart Scientific	5628	Temperature, Probe	7/28/2023	1/31/2025
CP32035	Hart Scientific	5628	Temperature, Probe	5/04/2023	11/30/2024
CP371282	Fluke Corporation	1594A	Precision Readout	5/27/2022	3/31/2025
CP50205	Hart Scientific	5628	Temperature, Probe	12/08/2023	6/30/2025

Calibration Data

Function Tested	Nominal / Reference Value	Measured Value	OOT	Calibration Tolerance <i>g: = Guard Banding Applied</i>	TUR	EMU
Temperature Accuracy	-15.000 °C					
As Found & As Left	-15.000	-14.99		-15.050 to -14.950 °C	3.3:1	± 0.015 °C
	0.000 °C					
As Found & As Left	0.000	0.01		-0.030 to 0.030 °C	2.4:1	± 0.013 °C
	75.000 °C					
As Found & As Left	75.000	75.00		74.970 to 75.030 °C	2.0:1	± 0.015 °C
	125.000 °C					
As Found & As Left	125.000	124.95		124.950 to 125.050 °C	3.0:1	± 0.016 °C

Temperature: 22 °C
Humidity: 38 %RH
Rpt. No.: 1756976

Calibration Performed By:		Quality Reviewer:	
Chris Lu	Technician, Metrology	James Alexander	5/31/2024
Name	Title	Name	Date

This report may not be reproduced, except in full, without written permission of Innocal. The results stated in this report relate only to the items tested or calibrated. Measurements reported herein are traceable to SI units via national standards maintained by NIST and were performed in compliance with MIL-STD-45662A, ANSI/NCSL Z540-1-1994, 10CFR50, Appendix B, ISO 9002-94, and ISO 17025:2017. Conformance based on Simple Acceptance as defined in ILAC G8 with a < 50% PFA. The estimated measurement uncertainty (EMU) if reported on this certificate, is being reported at a confidence level of 95% or K=2 unless otherwise noted in the remarks section.

*7-22-24

SERVICE AUTHORIZATION FORM
Highway Safety Technology Unit
80 South Swan Street
Albany, NY 12210
Ph: 518-485-7636 / 518-402-0689
Fax: 518-457-6869

THIS FORM IS USED TO AUTHORIZE THE DIVISION OF CRIMINAL JUSTICE SERVICES TO PERFORM SERVICE ON THE INSTRUMENT IDENTIFIED. THIS AUTHORIZATION MUST ACCOMPANY EACH PIECE OF EQUIPMENT WHEN IT IS PRESENTED TO THE EQUIPMENT REPAIR CENTER FOR SERVICE. **PLEASE SIGN AND DATE FORM.**

SECTION I - AGENCY INFORMATION

Name of Agency WARREN COUNTY SHERIFF'S OFFICE		Today's Date 7/22/24
Address 1400 STATE ROUTE 9		City, State, ZIP LAKE GEORGE, NY 12845
Contact Person PATROL SERGEANT BEN GEISLER		Contact Telephone (518) 743-2500
Contact Email ben.geisler@warrencountysheriffny.us		Best Way to Reach You During Appointment Period (518) 222-1089

SECTION II - INSTRUMENT INFORMATION

Instrument Manufacturer (Please Circle) NPAS DMT (GUTH) Applied Concepts (Stalker) Decatur (Genesis) Kustom (Eagle/Falcon) MPH Alco-Sensor Other:		
Model Number	Instrument Serial Number MJ 4824	
Antenna Serial Number (if applicable) Antenna 1 -- / Antenna 2 --	Type of maintenance (check appropriate) <input type="checkbox"/> Calibration Check <input type="checkbox"/> Repair	
Brief Description of Malfunction ANNUAL REIDENTIFICATION		
List all parts, cables and/or accessories being submitted. Tuning forks must accompany all radar units (if applicable). (Please Circle)		
DMT/SIM: DMT Keyboard	SIM (1)* Other: (1)* Power case Tubes (1) communication cable	Radar: Control Unit Antenna(s) Antenna Cable(s) Power Cord Forks Other:
I acknowledge I am signing this Service Authorization Form as an act and deed of said agency, organization, or municipality, and that I am duly authorized to sign same for the uses and purposes mentioned herein. PLEASE SIGN AND DATE.		
Signature PO [Signature] #427		Date 07/22/2024

DCJS USE ONLY

Export °C <u>34.00</u> <u>34.00</u> <u>34.00</u>	Digital °C <u>34.02</u> <u>34.02</u> <u>34.02</u>	Thermometer: <u>307443</u> Calibration Due: <u>5/31/2025</u> Air Leak Test: <u>OK</u> Propeller Rotation: <u>OK</u>
		* @ 7-22-24