



**CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS  
MAINTAINED IN THE REGULAR COURSE OF BUSINESS**

I, Steven J. Carluccio, Highway Safety Equipment Technical Supervisor, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of GUTH LABORATORIES, 12V500 SIMULATOR, serial number MP4824, performed on July 25, 2022 by Office of Public Safety employee Martin Schilling, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/ maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Steven J. Carluccio".

Digitally signed under ESRA  
by Steven J. Carluccio  
on 08/04/2022 12:50 PM

---

**Steven J. Carluccio  
Highway Safety Equipment Technical Supervisor  
Office of Public Safety**

**SIMULATOR  
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

**Name of Submitting Agency: Warren County Sheriff's Office**

**Manufacturer: GUTH LABORATORIES**

**Model: 12V500**

**Serial Number: MP4824**

**Date: July 25, 2022**

<b>Breath Simulator Temperature</b>
34.01 degrees centigrade



**Division of Criminal  
Justice Services**

**I hereby certify that I have performed any necessary maintenance procedures and verified the calibration of GUTH LABORATORIES simulator model 12V500, serial number MP4824, and have determined that it is accurate.**

**I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified simulator was performed, or within a reasonable time thereafter.**

A handwritten signature in black ink, appearing to read "Martin Schilling".

Digitally signed under ESRA  
by Martin Schilling  
on 07/25/2022 01:40 PM

---

**Martin Schilling  
Highway Safety Equipment Technician  
Office of Public Safety**

**SIMULATOR  
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

**Name of Submitting Agency: Warren County Sheriff's Office**

**Manufacturer: GUTH LABORATORIES**

**Model: 12V500**

**Serial Number: MP4824**

**Date: July 25, 2022**

**1. General appearance and condition of this device -----: OK**

**2. Check propeller rotation -----: OK**

**3. Check for air leaks -----: OK**

**4. Check for proper temperature -----: OK**

**5. Record any repairs made to this instrument and list services provided : VERIFIED  
CALIBRATION**



**Division of Criminal  
Justice Services**

A handwritten signature in black ink that reads "Martin Schilling".

Digitally signed under ESRA  
by Martin Schilling  
on 07/25/2022 01:40 PM

---

**Martin Schilling  
Highway Safety Equipment Technician  
Office of Public Safety**

## NIST Traceable Calibration Report



1681893

Reference Number: 1416678  
 PO Number: SCARLUCCIO030222

New York State Division Criminal Justice Services  
 80 South Swan Street  
 Albany, NY 12210 United States

<b>Manufacturer:</b> Alpha Technics	<b>Calibration Date:</b> 03/11/2022
<b>Model Number:</b> 91-5443	<b>Calibration Due Date:</b> 03/11/2023
<b>Description:</b> Temperature, Thermometer	<b>Condition As Found:</b> In Tolerance
<b>Asset Number:</b> 20BMM01712	<b>Condition As Left:</b> In Tolerance, No adjustment
<b>Serial Number:</b> 20BMM01712	
<b>Procedure:</b> DS Alpha Technics 5000 Series f/NY State DCJS	

**Remarks:**  
 NIST-traceable calibration performed on the unit referenced above in accordance with customer requirements, published specifications and the lab's standard operating procedures. No adjustments were made to the unit.

### Standards Utilized

Asset No.	Manufacturer	Model No.	Description	Cal. Date	Due Date
CP144688	Fluke Corporation	5628	Temperature, PRT Probe	07/15/2021	07/31/2022
CP144944	Fluke Corporation	1523	Temperature, Indicator	05/28/2021	05/31/2022

### Calibration Data

FUNCTION TESTED	Nominal Value	As Found	Out of Tol	As Left	Out of Tol	CALIBRATION TOLERANCE
Temperature	-15.000 °C	-14.96		Same		-15.050 to -14.950 °C [EMU 0.017 °C][TUR 2.9:1]
	0.000 °C	0.01		Same		-0.030 to 0.030 °C [EMU 0.017 °C][TUR 1.8:1]
	75.000 °C	75.01		Same		74.970 to 75.030 °C [EMU 0.022 °C][TUR 1.4:1]
	125.000 °C	125.03		Same		124.950 to 125.050 °C [EMU 0.025 °C][TUR 2.0:1]

**Temperature:** 25° C  
**Humidity:** 30% RH  
**Rpt. No.:** 1681893

Calibration Performed By:				Quality Reviewer:	
Name	ID #	Title	Phone	Name	Date
Bautista, Arnel	358	Metrologist	847-327-5339	Szplit, Tony	03/11/2022

This report may not be reproduced, except in full, without written permission of Innocal. The results stated in this report relate only to the items tested or calibrated. Measurements reported herein are traceable to SI units via national standards maintained by NIST and were performed in compliance with MIL-STD-45662A, ANSI/NCSL Z540-1-1994, 10CFR50, Appendix B, ISO 9002:94, and ISO 17025:2017. Conformance based on Simple Acceptance as a Decision Rule. The estimated measurement uncertainty (EMU), if reported on this certificate, is being reported at a confidence level of 95% or K=2 unless otherwise noted in the remarks section.

**SERVICE AUTHORIZATION FORM**  
**Highway Safety Technology Unit**  
 80 South Swan Street  
 Albany, NY 12210  
 Ph: 518-485-7636 / 518-402-0689  
 Fax: 518-457-6869

*MJS  
7/25/22*

THIS FORM IS USED TO AUTHORIZE THE DIVISION OF CRIMINAL JUSTICE SERVICES TO PERFORM SERVICE ON THE INSTRUMENT IDENTIFIED. THIS AUTHORIZATION MUST ACCOMPANY EACH PIECE OF EQUIPMENT WHEN IT IS PRESENTED TO THE EQUIPMENT REPAIR CENTER FOR SERVICE. **PLEASE SIGN AND DATE FORM.**

**SECTION I - AGENCY INFORMATION**

Name of Agency <b>WARREN COUNTY SHERIFF'S OFFICE</b>		Today's Date <b>7/25/22</b>
Address <b>1400 STATE ROUTE 9</b>		City, State, ZIP <b>LAKE GEORGE, NY 12845</b>
Contact Person <b>SGT. C.B. GEISLER</b>		Contact Telephone <b>518-743-2500</b>
Contact Email <b>BEN.GEISLER@WARRENCOUNTYSHERIFFNY.U</b>		Best Way to Reach You During Appointment Period <b>518-470-5219</b>

**SECTION II - INSTRUMENT INFORMATION**

Instrument Manufacturer (Please Circle) NPAS DMT <u>(GUTH)</u> Applied Concepts (Stalker) Decatur (Genesis) Kustom (Eagle/Falcon) MPH Alco-Sensor Other:			
Model Number <b>12V500</b>		Instrument Serial Number <b>MP4824</b>	
Antenna Serial Number (if applicable) Antenna 1 -- / Antenna 2 --			Type of maintenance (check appropriate) <input checked="" type="checkbox"/> Calibration Check <input type="checkbox"/> Repair
Brief Description of Malfunction			
List all parts, cables and/or accessories being submitted. Tuning forks must accompany all radar units (if applicable). (Please Circle)			
DMT/SIM: DMT                      SIM                      Tubes Keyboard                Other:		Radar: Control Unit            Antenna(s)            Antenna Cable(s) Power Cord             Forks                    Other:	
I acknowledge I am signing this Service Authorization Form as an act and deed of said agency, organization, or municipality, and that I am duly authorized to sign same for the uses and purposes mentioned herein. <b>PLEASE SIGN AND DATE.</b>			
Signature <i>[Signature]</i>		Date <b>07/25/2022</b>	

**DCJS USE ONLY**

Export °C <u>34.00</u> <u>34.00</u> <u>34.00</u>	Digital °C <u>24.01</u> <u>24.01</u> <u>24.01</u>	Thermometer: <u>20 BMM 01712</u> Calibration Due: <u>3-11-23</u> Air Leak Test: <u>06</u> Propeller Rotation: <u>06</u>
---	--	---