



**CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS
MAINTAINED IN THE REGULAR COURSE OF BUSINESS**

I, Steven J. Carluccio, Highway Safety Equipment Technical Supervisor, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of GUTH LABORATORIES, 12V500 SIMULATOR, serial number MP4824, performed on December 06, 2022 by Office of Public Safety employee Martin Schilling, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/ maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Steven J. Carluccio".

Digitally signed under ESRA
by Steven J. Carluccio
on 12/14/2022 10:05 AM

**Steven J. Carluccio
Highway Safety Equipment Technical Supervisor
Office of Public Safety**

**SIMULATOR
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Warren County Sheriff's Office

Manufacturer: GUTH LABORATORIES

Model: 12V500

Serial Number: MP4824

Date: December 06, 2022

Breath Simulator Temperature
34.01 degrees centigrade



**Division of Criminal
Justice Services**

I hereby certify that I have performed any necessary maintenance procedures and verified the calibration of GUTH LABORATORIES simulator model 12V500, serial number MP4824, and have determined that it is accurate.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified simulator was performed, or within a reasonable time thereafter.

A handwritten signature in black ink that reads "Martin Schilling".

Digitally signed under ESRA
by Martin Schilling
on 12/06/2022 08:32 AM

**Martin Schilling
Highway Safety Equipment Technician
Office of Public Safety**

**SIMULATOR
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Warren County Sheriff's Office

Manufacturer: GUTH LABORATORIES

Model: 12V500

Serial Number: MP4824

Date: December 06, 2022

1. General appearance and condition of this device -----: OK

2. Check propeller rotation -----: OK

3. Check for air leaks -----: OK

4. Check for proper temperature -----: OK

**5. Record any repairs made to this instrument and list services provided : VERIFIED
CALIBRATION**



**Division of Criminal
Justice Services**

A handwritten signature in black ink that reads "Martin Schilling".

Digitally signed under ESRA
by Martin Schilling
on 12/06/2022 08:32 AM

**Martin Schilling
Highway Safety Equipment Technician
Office of Public Safety**

NIST Traceable Calibration Report



1681893

Reference Number: 1416678

PO Number: SCARLUCCIO030222

New York State Division Criminal Justice Services
 80 South Swan Street
 Albany, NY 12210 United States

Manufacturer: Alpha Technics	Calibration Date: 03/11/2022
Model Number: 91-5443	Calibration Due Date: 03/11/2023
Description: Temperature, Thermometer	Condition As Found: In Tolerance
Asset Number: 20BMM01712	Condition As Left: In Tolerance, No adjustment
Serial Number: 20BMM01712	
Procedure: DS Alpha Technics 5000 Series f/NY State DCJS	

Remarks:

NIST-traceable calibration performed on the unit referenced above in accordance with customer requirements, published specifications and the lab's standard operating procedures. No adjustments were made to the unit.

Standards Utilized

Asset No.	Manufacturer	Model No.	Description	Cal. Date	Due Date
CP144688	Fluke Corporation	5628	Temperature, PRT Probe	07/15/2021	07/31/2022
CP144944	Fluke Corporation	1523	Temperature, Indicator	05/28/2021	05/31/2022

Calibration Data

FUNCTION TESTED	Nominal Value	As Found	Out of Tol	As Left	Out of Tol	CALIBRATION TOLERANCE
Temperature	-15.000 °C	-14.96		Same		-15.050 to -14.950 °C [EMU 0.017 °C][TUR 2.9:1]
	0.000 °C	0.01		Same		-0.030 to 0.030 °C [EMU 0.017 °C][TUR 1.8:1]
	75.000 °C	75.01		Same		74.970 to 75.030 °C [EMU 0.022 °C][TUR 1.4:1]
	125.000 °C	125.03		Same		124.950 to 125.050 °C [EMU 0.025 °C][TUR 2.0:1]

Temperature: 25° C
Humidity: 30% RH
Rpt. No.: 1681893

Calibration Performed By:				Quality Reviewer:	
Name	ID #	Title	Phone	Name	Date
Bautista, Arnel	358	Metrologist	847-327-5339	Szplit, Tony	03/11/2022

This report may not be reproduced, except in full, without written permission of Innocal. The results stated in this report relate only to the items tested or calibrated. Measurements reported herein are traceable to SI units via national standards maintained by NIST and were performed in compliance with MIL-STD-45662A, ANSI/NCSL Z540-1-1994, 10CFR50, Appendix B, ISO 9002:94, and ISO 17025:2017. Conformance based on Simple Acceptance as a Decision Rule. The estimated measurement uncertainty (EMU), if reported on this certificate, is being reported at a confidence level of 95% or K=2 unless otherwise noted in the remarks section.

