

CERTIFICATE OF PHOTOSTATIC COPY OF
RECORD OF ANALYSIS -- SIMULATOR SOLUTION
(CPLR RULE 4518)

I, GERALD M. ZEOSKY, Inspector, Director of the New York State Police Crime Laboratories, Forensic Investigation Center, State Campus Building #30, 1220 Washington Avenue, Albany, New York 12226-3000, do hereby certify and authenticate, as provided by Rule 4518, Subdivision c, Civil Practice Law and Rules, that the copy annexed hereto is an exact photocopy of the original record of the New York State Police Headquarters Crime Laboratory which has been delegated to my possession, custody and control, by the Superintendent of the New York State Police.

I further certify that the original Record of Analysis -- Simulator Solution Lot Number *09060*, performed on *March 19, 2009* by State Police Crime Laboratory employee *Harry K. Garber, Forensic Scientist III*, of which the copy annexed is a photocopy, was made in the regular course of business of the New York State Police Crime Laboratory, that it is the regular course of Crime Laboratory business to make such record and that such record was made at the time of each event recorded in it or within a reasonable time thereafter.

April 27, 2009


Inspector Gerald M. Zeosky
Director



HARRY J. CORBITT
SUPERINTENDENT

NEW YORK STATE POLICE
FORENSIC INVESTIGATION CENTER
Building 30
1220 Washington Avenue
Albany, New York 12226-3000
(518) 457-1208

March 19, 2009

CERTIFICATION OF ANALYSIS
0.10% BREATH ALCOHOL SIMULATOR SOLUTION

Date of Analysis: March 19, 2009
Supplied By: Guth Laboratories
Lot Number: 09060
Expiration Date: September 2, 2009
Method Of Analysis: Headspace Gas Chromatography
Results of Analysis: 123.5 mg Ethyl Alcohol / 100 mL of solution

Simulator solution lot number 09060 has been certified to contain the appropriate concentration of ethyl alcohol and is hereby approved for use.

When this reference solution is used with a properly operating breath testing instrument,
the solution will provide a value of 0.10% at 34°C.

Harry K. Garber

Forensic Scientist III
Toxicology

Keith K. Coonrod

Director of Toxicology and Drug Chemistry
Forensic Investigation Center