

CERTIFIED COPY (CPLR RULE 4518)

**0.10% SIMULATOR SOLUTION RECORD**

Lot Number 11010, dated February 10, 2011

I, Jennifer F Limoges, Supervisor of Forensic Services, Forensic Investigation Center, do hereby certify and authenticate, as provided by Rule 4518, subdivision c, Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the New York State Police Forensic Investigation Center which has been delegated to my possession, custody and control by the Superintendent of the New York State Police. This record was made in the regular course of business of the New York State Police Forensic Investigation Center, it is the regular course of the Forensic Investigation Center business to make such record, and such record was made at the time of each event recorded in it or within a reasonable time thereafter.

 Digitally signed under  
ESRA by Jennifer F  
Limoges on  
2011.02.18 07:09:07 -05'00'

Jennifer F Limoges

Supervisor of Forensic Services  
Forensic Investigation Center



JOSEPH A. D'AMICO  
SUPERINTENDENT

**NEW YORK STATE POLICE  
FORENSIC INVESTIGATION CENTER  
BUILDING 30  
1220 WASHINGTON AVE.  
ALBANY, NY 12226-3000**

**0.10% SIMULATOR SOLUTION RECORD**

**Report Date:** February 10, 2011  
**Manufacturer:** Guth Laboratories, Inc.  
**Lot Number:** 11010  
**Expiration Date:** 7/4/2011  
**Analysis Results:** 0.120 g/dL

Simulator solution lot number 11010 has been tested by headspace gas chromatography to contain the appropriate concentration of ethyl alcohol and is hereby approved for use.

When this simulator solution is used with a properly operating breath testing instrument, it will provide a value of 0.10% within acceptable limits.

Digitally signed under  
ESRA by Carrie A Kirkton  
on  
2011.02.10 10:00:27 -05'00'

Carrie A Kirkton

Forensic Scientist III  
Breath Analysis Technical Supervisor

Digitally signed under  
ESRA by Jennifer F  
Limoges on  
2011.02.18 07:08:31 -05'00'

Jennifer F Limoges

Supervisor of Forensic Services  
Breath Analysis Technical Supervisor